



# Reagecon

A CALIBRE SCIENTIFIC COMPANY

January 2025

To whom it may concern,

Reagecon Standards should be stored in their original container (in an upright position) at the applicable temperature outlined on the Product / Box Label ( $< -15^{\circ}\text{C}$  or  $2^{\circ}\text{C}$  to  $8^{\circ}\text{C}$  or  $5^{\circ}\text{C}$  to  $30^{\circ}\text{C}$ ), away from any strong light source (e.g. sunshine, UV lamp), heat sources and volatile chemicals. It is strongly advised to record the first opening of the bottle, to close it immediately after every use and to discard the contents of the bottle when less than 10% of the original volume of the solution is left. The expiry date as outlined on the bottle label is valid provided the product is properly stored and handled in compliance with Good Laboratory Practice, with the exception of the following parts which are intended for single use:

- Conductivity Standard  $1.3\ \mu\text{S}/\text{cm}$
- Premium TOC range (single use vial)
- PAHs / Phenols / Pesticides / VOCs / PCBs / Azo Dyes / Arocolors / Toxaphene / Osmolality (in ampoules).

The certification ceases to be valid if the product is damaged, contaminated or modified.

Regards  
Darren McGrath  
Quality Director

Quality Department  
Reagecon

If you have any questions regarding the above, please do not hesitate to contact us at [quality@reagecon.ie](mailto:quality@reagecon.ie)